

Mammography QC Package

Insure your images are at their very best

The CIRS Model 029-15 Mammography Quality Control Package has the tools a technologist needs to implement a comprehensive quality control program and insure their department is producing the best possible images and diagnostic information.

Standard tests include evaluation of processor performance, fixer retention, image quality, film screen contact and paddle compression. The package includes items necessary for implementing a Mammography QC program in compliance with FDA / MQSA regulations, ACR recommendations and state requirements.



Model 029-15

Set Includes

- Lint free wipes
- Digital thermometer
- Analog Scale
- Hypo test kit
- Film contact test tool
- Phantom Model 015
- Sensitometer
- Densitometer
- High impact carry case

Optional Accessories

Program Manual for accreditation program sponsored by the American College of Radiology

*Tissue Simulation &
Phantom Technology*

This product is available through:

JRT Associates

800-221-0111

Phantom Specifications

WAX INSERT

FIBERS


1. 1.56 mm nylon fiber
2. 1.12 mm nylon fiber
3. 0.89 mm nylon fiber
4. 0.75 mm nylon fiber
5. 0.54 mm nylon fiber
6. 0.40 mm nylon fiber

SPECKS

7. 0.54 mm Al_2O_3 speck
8. 0.40 mm Al_2O_3 speck
9. 0.32 mm Al_2O_3 speck
10. 0.24 mm Al_2O_3 speck
11. 0.16 mm Al_2O_3 speck

MASSES

12. 2.00 mm (thickness) mass
13. 1.00 mm (thickness) mass
14. 0.75 mm (thickness) mass
15. 0.50 mm (thickness) mass
16. 0.25 mm (thickness) mass



PHANTOM BODY

MATERIAL: ACRYLIC
 LENGTH: 10.8 CM
 WIDTH: 10.15 CM
 DEPTH: 4.4 CM

